



PRODUCT BULLETIN

Product(s):	Narkomed® MRI-2
Bulletin Number:	140
Release Date:	September 3, 2002
Pages:	1

Subject: 3 T (Tesla) Documentation

The FDA has granted 510(k) clearance allowing unrestricted operation in MRI environments of 3 Tesla or less for all Narkomed MRI-2 anesthesia machines. Also included in the clearance is an MRI Bain circuit, which can be used with any Narkomed MRI-2 machine.



A kit is available which provides a label for the reverse side of the MRI-2 with the 3 Tesla limitations, and a new Operator's Manual that includes Bain circuit operation.

Part Number 4118006

MRI-2 3T Upgrade Kit

\$450

If you have any questions or need information, call the Product Manager, Greg Sutherland at extension #2311.